



### SERIES 2 — Standard Type, 0.001 mm & 0.005 mm Graduation

- Standard 0.001 mm and 0.005 mm graduation dial indicators having a bezel with an outside diameter of 57 mm.
- The bezel clamp and lifting lever (optional) can be attached to either the right or left side. These parts can be easily installed and removed without tools.
- Watertight assembly of bezel and crystal as well as the use of an O-ring prevents water or oil penetration.
- The stem and spindle are made of high-strength quench-hardened stainless steel suitable for heavy-duty use.
- A carbide contact point is used.
- A special alloy is used for the sector gears to provide improved wear resistance.
- The indicator uses jeweled bearings, providing excellent indication sensitivity and durability.
- Application of an anti-reflective and hard surface coating improves scale visibility along with scratch and chemical resistance.










## ISO/JIS Type



Order No.	А	В	С	D	E	F	G	Н	I	Mass (g) w/lug   Flat-back	
2109A-10	47.2	60.5	57	18.1	20	16.9	15.1	52	7.6	148	139
2110A-10	47.2	66.5	57	18.1	20	16.9	21.1	52	7.6	149	140
2113A-10	47.2	61	57	18.1	20	16.9	15.6	52	7.6	148	139
2118A-10	47.2	60.7	57	18.1	20	16.9	15.3	52	7.6	146	137
2119A-10	47.2	60.7	57	18.1	20	16.9	15.3	52	7.6	146	137
2124A-10	47.2	60.7	57	18.1	20	16.9	15.3	52	7.6	146	137

Note: Refer to pages F-57 to F-60 for details of contact points.

# Optional Accessories • Limit hand (2 pcs.): 21AZB195





• Bezel clamp: 21AZB148



#### **FEATURES**

Metric							
Ord	Order No.				(a)	$\square$	兩
w/lug	Flat-back	90 0 10	10 0 10	9	64	$\square$	
2109A-10	2109AB-10		~	~		~	
2110A-10	2110AB-10	~		~		~	~
2113A-10	2113AB-10		~	~		~	
2118A-10	2118AB-10	~				~	
2119A-10	2119AB-10		~			~	
2124A-10	2124AB-10	~				~	

### **SPECIFICATIONS**

Metric							ISO/JIS type							
Order No.		Graduation	Range		Maximum	permissil	ole error (I	MPE) (μm)		Dial	Manaurina			
	w/lug	Flat-back	(mm)		(range/rev)	Indication error			Hysteresis	Repeat-	reading	Measuring force (N)		
	writing	I lat-back		(mm)	1/10 Rev	1/2 Rev	1 Rev	Measuring range	TTYSTETESIS	ability	reduing	10100 (14)		
	2109A-10	2109AB-10	0.001	1 (0.2)	2	3	4	5	2	0.5	0-100-0	1.5 or less		
	2110A-10	2110AB-10	0.001	1 (0.1)	2	3	4	5	2	0.5	±0-100	1.8 or less		
	2113A-10	2113AB-10	0.001	2 (0.2)	2	4	5	7	2	0.5	0-100-0	1.5 or less		
	2118A-10	2118AB-10	0.001	5 (0.2)	3.5	5	6	10	3	1	0-100-100	1.5 or less		
	2119A-10	2119AB-10	0.001	5 (0.2)	3.5	5	6	10	3	1	0-100-0	1.5 or less		
	2124A-10	2124AB-10	0.005	5 (0.5)	5	8	9	12	3	3	±0-50	1.5 or less		

Note: Completed products inspection is performed in the vertical orientation (contact point downward) and the stated accuracy is guaranteed.